

## Bemerkungsfeld in den Röhrendaten

The screenshot shows the 'Röhrenname' field containing 'EAM86'. The 'System 1' dropdown is set to 'Mag.Auge+', 'System 2' to 'Diode', and 'System 3' to 'Triode'. The 'Sockel/Fassung' section includes a diagram of a 9-pin tube base and a photograph of a 'Noval B9A' tube. The 'Bemerkungen zur Röhre' field is highlighted with a red border and contains the following text:

**Bemerkungen zur Röhre:** Hilfe zu Röhrenart:  
 = 6GX8  
 !Mag.Auge / Magic Eye: Test: Ust +35 ... +250V;  
 Steuerung über Steuersteg mit positiver Spannung. Bei etwa Ust=165 sollten sich die Leucht balken berühren.

Im Bemerkungsfeld können Angaben zur Röhre gemacht werden, für welche kein eigenes Feld vorgesehen ist. Die Angaben sollen nach folgenden Regeln erfolgen:

1. An erster Stelle stehen Vergleichsröhren, Beispiel: = EF80, AZ12, EM34
2. es können beliebige Angaben gemacht werden. Dabei gibt es folgende Kürzel (Beispiel):

V: Senderöhre - Verwendungszweck

3. An letzter Stelle folgt Instruction/Info. Alles was nach einem ‚!‘ folgt, wird im manuellen Modus und manuellen Modus mit Vorwiderstand angezeigt